

# checkCIF/PLATON report

You have not supplied any structure factors. As a result the full set of tests cannot be run.

No syntax errors found.      CIF dictionary      Interpreting this report

## Datablock: jad693si

---

Bond precision:	C-C = 0.0030 A	Wavelength=0.71073	
Cell:	a=16.5114(7)	b=5.3548(2)	c=16.3972(7)
	alpha=90	beta=109.139(2)	gamma=90
Temperature:	100 K		
	Calculated	Reported	
Volume	1369.63(10)	1369.63(10)	
Space group	P 21/c	P 2(1)/c	
Hall group	-P 2ybc	?	
Moiety formula	C24 H36 N2 O8 Pd, 2(H2 O)	?	
Sum formula	C24 H40 N2 O10 Pd	C24 H40 N2 O10 Pd	
Mr	622.98	622.98	
Dx,g cm-3	1.511	1.511	
Z	2	2	
Mu (mm-1)	0.734	0.734	
F000	648.0	648.0	
F000'	646.33		
h,k,lmax	21,7,21	21,7,21	
Nref	3334	3168	
Tmin,Tmax	0.853,0.936	0.838,0.937	
Tmin'	0.832		

Correction method= MULTI-SCAN

Data completeness= 0.950      Theta(max)= 28.130

R(reflections)= 0.0301( 2979)      wR2(reflections)= 0.0671( 3168)

S = 1.117      Npar= 188

---

The following ALERTS were generated. Each ALERT has the format

**test-name\_ALERT\_alert-type\_alert-level.**

Click on the hyperlinks for more details of the test.

---

### Alert level C

PLAT094\_ALERT\_2\_C Ratio of Maximum / Minimum Residual Density .... 2.57

---

### Alert level G

```

HYDTR01_ALERT_1_G Extra text has been found in the _refine_ls_hydrogen_treatment fi
    Explanatory text should be in the _publ_section_refinement field.
    Hydrogen treatment given as water H free with SADI, NH2 free, rigid
    Hydrogen treatment identified as riding
PLAT002_ALERT_2_G Number of Distance or Angle Restraints on AtSite          3
PLAT005_ALERT_5_G No _iucr_refine_instructions_details in CIF ....         ?
PLAT720_ALERT_4_G Number of Unusual/Non-Standard Labels .....             4
PLAT860_ALERT_3_G Note: Number of Least-Squares Restraints .....          1

```

---

```

0 ALERT level A = Most likely a serious problem - resolve or explain
0 ALERT level B = A potentially serious problem, consider carefully
1 ALERT level C = Check. Ensure it is not caused by an omission or oversight
5 ALERT level G = General information/check it is not something unexpected

```

```

1 ALERT type 1 CIF construction/syntax error, inconsistent or missing data
2 ALERT type 2 Indicator that the structure model may be wrong or deficient
1 ALERT type 3 Indicator that the structure quality may be low
1 ALERT type 4 Improvement, methodology, query or suggestion
1 ALERT type 5 Informative message, check

```

---

It is advisable to attempt to resolve as many as possible of the alerts in all categories. Often the minor alerts point to easily fixed oversights, errors and omissions in your CIF or refinement strategy, so attention to these fine details can be worthwhile. In order to resolve some of the more serious problems it may be necessary to carry out additional measurements or structure refinements. However, the purpose of your study may justify the reported deviations and the more serious of these should normally be commented upon in the discussion or experimental section of a paper or in the "special\_details" fields of the CIF. checkCIF was carefully designed to identify outliers and unusual parameters, but every test has its limitations and alerts that are not important in a particular case may appear. Conversely, the absence of alerts does not guarantee there are no aspects of the results needing attention. It is up to the individual to critically assess their own results and, if necessary, seek expert advice.

### **Publication of your CIF in IUCr journals**

A basic structural check has been run on your CIF. These basic checks will be run on all CIFs submitted for publication in IUCr journals (*Acta Crystallographica*, *Journal of Applied Crystallography*, *Journal of Synchrotron Radiation*); however, if you intend to submit to *Acta Crystallographica Section C* or *E*, you should make sure that full publication checks are run on the final version of your CIF prior to submission.

### **Publication of your CIF in other journals**

Please refer to the *Notes for Authors* of the relevant journal for any special instructions relating to CIF submission.

---

**PLATON version of 18/07/2011; check.def file version of 04/07/2011**

